

- 2) Liveon Biolbs Pvt. Ltd., plot no. 46 & 47, II Phase, Water Tank Road, Kiadb Industrial Area, Antharasanahalli, Tumkur, Karnataka
 Contact Details: jyothi@liveonbiolabs.com Ms. Jyoti 9480849735

| S. No. | Product | Standard | Test |
|--------|--|---|--|
| 1 | Gloves, catheters, bronchoscope, solution & blood administration set, transfusion set, drainage set, oxygenator, extracorporeal oxygenator tubing and accessories, drug delivery catheters, sutures, needles | ISO 10993-5 ISO 10993-3 ISO 10993-10 ISO 10933-4 ISO 10993-11 | Cytotoxicity Test, Elution method, Agar Overlay Assay, Direct Contact Assay, NRU Cytotoxicity, MTT Cytotoxicity Genotoxicity Test, In Vitro -CHO Cell lines/In Vivo Chromosomal Aberration Test, Bacterial Reverse Mutation Test (Ames Test) In Vitro Chromosomal Aberration Test in Human Peripheral Blood Lymphocytes, In Vitro -CHO Cell lines/In Vivo Micronucleus Test, In Vitro Mammalian Cell Gene Mutation Test Using the Thymidine Kinase Gene Irritation Test , Primary Skin irritation Test and Intracutaneous Reactivity Test, Skin Sensitization Test - Guinea Pig Maximization Test, Buehler Test and Local Lymph Node Assay Interaction with blood Bacterial endotoxin test |

- 3) Radiant Research Services Private Limited, 99/A, 8th Main, 3rd Phase, Peenya Industrial Area, Bangalore
 Contact Details: qms@radiantresearch.in Dr. Manjunath S Ph. 9449169172

| S. No. | Product | Standard | Test |
|--------|--|---|---|
| 1 | Medical Devices (Disposable Perfusion Sets, I.V. Cannula, Blood bags, Wound Suture And Closure Products, Peritoneal Dialysis Sets, Scalp Vein, Infusion tube, Stents, Catheters, Intra Ocular Lenses, Gels, Heart Valves, Gloves, Wound Dressing Materials, Intra-vaginal and inta-intestinal devices, Orthodontic Devices, Dental Prosthesis, Dental Filling Materials) | ISO 10993 -11 ISO 10993- 10 ISO 10993 – 11 & ISO 10993 – 12 ISO 10993 – 05 & ISO 10993- 12 | Acute systemic toxicity Skin irritation, eye irritation, skin sensitization Material mediated pyrogenicity Cytotoxicity test, NRU cytotoxicity, MTT cytotoxicity |

STATE: DELHI

- 4) Shriram Institute for Industrial Research (A Unit of Shriram Scientific & Industrial Research Foundation), 19, University Road, Delhi
Contact Details: laxmirawat@shriraminstitute.org Ph. 9818193810/08447705730 Dr. Laxmi Rawat

| S. No. | Product | Standard | Test |
|---------------|--|--|--|
| 1 | Surface Device Skin Electrodes, External Protheses, Fixation Tapes, Compression Bandages, Monitors of Various Types | ISO-10993-5 ISO 10993- 10 | Cytotoxicity Irritation / Intracutaneous reactivity |
| 2 | Mucosal Membrane Contact Devices, Endotracheal Tubes, Bronchoscopes | ISO 10993-5 ISO 10993-10 ISO 10993-3 | Cytotoxicity Irritation, Intracutaneous reactivity Genotoxicity |
| 3 | Blood Path, Indirect: Solution Administration Sets, Extension Sets, Transfer Sets, Blood Admin Sets | ISO 10993-5 ISO 10993-10 ISO 10993-11,12 ISO 10993-3 ISO 10993-4 | Cytotoxicity Irritation, Intracutaneous reactivity Systemic toxicity(Acute/short term) Carcinogenicity, Genotoxicity Haemolytic Properties of Material Haemo-compatibility test |
| 4 | Circulating Blood: Intravascular catheters, Temporary pacemaker electrodes, Oxygenators, Extracorporal oxygenator Tubing and accessories | ISO 10993-5 ISO 10993-10 ISO 10993-11,12 ISO 10993-3 ISO 10993-4 | Cytotoxicity Irritation, Intracutaneous reactivity Systemic toxicity(Acute/short term) Carcinogenicity, Genotoxicity Haemolytic Properties of Material Haemo-compatibility test |

STATE: TAMIL NADU

- 5) Vanta Bioscience Limited, K2b, 11th Cross Street, Sipcot Industrial Complex, Gummidipoondi, Thiruvallur, Tamil Nadu
Contact Details: deepak.sharma@vantabio.com Mr. Deepak Sharma 9810299809

| S. No. | Product | Standard | Test |
|---------------|--|-----------------|--|
| 1 | Elbow Length Gloves, Powder free(PC), powder free sterile synthetic surgical gloves, powder free sterile surgical (micro) gloves, catheters, | ISO 10993 | Skin sensitization, Cytotoxicity Test: Elution method, Agar overlay assay, Direct contact assay, NRU Cytotoxicity, irritation test |

- 6) GLR Laboratories Private Limited, 444 Gokulam Street, Mathur, Chennai
Contact Details: Info@Glrilabs.Com Mr.K R Navaneethakrishnan Ph. 9841212596

| S. No. | Product | Standard | Test |
|---------------|---|--|--|
| 1 | Gloves, Respiratory devices, tubes, catheters, bags | ISO 10993-10 ISO 10993-4 ISO 10993-11 ISO 10993-5 | Intracutaneous test, skin irritation test, skin sensitization test Hemocompatibility, Hemolysis Acute Systemic toxicity, pyrogen Cytotoxicity |

STATE: TELANGANA

- 7) Palamur Biosciences Private Limited, Sh-20, Karvina, Madigattla Village, Bhoothpur Mandal, Mahbubnagar, Telangana
Contact Details: 9492352676 vd.giridharrao@palamurbio.com Mr. V D Giridhar Rao

| S. No. | Product | Standard | Test |
|---------------|--|--|--|
| 1 | Medical Devices (Surgical Gloves, Syringes with needles, Tubular Rings, Sutures, Surface devices, Catheters, External communicating devices) | ISO 10993-11 ISO 10993-6 ISO 10993-5 ISO 10993-10 | Acute Systemic Toxicity, Pyrogenicity test Implantation Tests In vitro Cytotoxicity Skin sensitization test, Irritation (eye/ocular, skin), intracutaneous reactivity |

STATE: GUJARAT

- 8) Accuprec Research Labs Pvt. Ltd., Opp. Zydus Pharmez, Changodar - Bavla Highway,, Ahmedabad
Contact Details : Rina.Gokani@Accuprec.Com Ms. Rina H. Gokani 9429571239

| S. No. | Product | Standard | Test |
|---------------|---|--|---|
| 1 | Surgical gloves, Catheters, syringes with needles, Tubular rings, sutures, Intraocular implants, vascular devices and IV sets | ISO 10993-4 EN ISO 10993-11 ISO 10993-3 ISO 10993-5 ISO 10993-10 | Hemocompatibility test, hemolysis test Pyrogen test, acute systemic toxicity Chromosomal aberration test, Ame's test, Micronucleus test Cytotoxicity MTT assay Eye irritation test, intracutaneous mucous membrane irritation test, skin sensitization test |

STATE: UTTAR PRADESH

- 9) Institute for Industrial Research & Toxicology, F-209, Phase-1, U.P.S.I.D.C., Masuri Dhaulna, Gulawati Marg, Ghaziabad, Uttar Pradesh.
Contact Details: Ms.Shalini Mishra, iirtdelhi@gmail.com 9711623081

| S. No. | Product | Standard | Test |
|---------------|--|------------------------------|---|
| 1 | Absorbent Cotton (Wound Surgical) | ISO 10993 | Mucous membrane irritation test on Rabbit, Skin sensitization test on Guinea Pig, |
| | Acetabular Shell (Orthopedics) | ISO 10993 | Mucous membrane irritation test on Rabbit, Skin sensitization test on Guinea Pig, Eye Irritation Test on Rabbit, Acute Toxicity dermal test on Mice, Acute Toxicity Dermal test on Rat, Acute Toxicity oral Test on Rat, Skin sensitization test on Guinea Pig, |
| | Bandages (Wound Surgical) Dialysis Tubing & Accessories (External Communicating device) | ISO 10993 | Skin sensitization test on Guinea Pig, |
| | Face Mask (Surface Device) | ISO 10993 | Eye irritation test on Rabbit |
| | Medical Device Invo Vac Foam Dressing | ISO 10993-10 ISO 10993-11 | Acute Dermal Toxicity (RAT) Acute Oral Toxicity (Rat) |
| | Wound Protector (Wound Surgical) | ISO 10993 | Skin sensitization test on Guinea Pig |